

Product Specification Sheet

Disposable Medical Gowns



Disposable Medical Gowns

Product Description

Disposable full-back fabric medical gowns protect health care providers against infectious disease. Gowns are examples of personal protective equipment used in health care settings. They are used to protect the wearer from the spread of infection or illness if the wearer comes in contact with potentially infectious liquid and solid material. They may also be used to help prevent the gown wearer from contaminating vulnerable patients, such as those with weakened immune systems

Features:

- Product Specification:
- This gown had reinforced front and arms
- High Quality Medical Gowns
- 10 Gowns per Bag
- High Quality AAMI Tested Fabric
- 2" White Elasticized Cuffs
- Full Back
- One Size Fits All

Gown Levels:

Level 1: Minimal risk, to be used, for example, during basic care, standard isolation, cover gown for visitors, or in a standard medical unit

Level 2: Low risk, to be used, for example, during blood draw, suturing, in the Intensive Care Unit (ICU), or a pathology lab

Level 3: Moderate risk, to be used, for example, during arterial blood draw, inserting an Intravenous (IV) line, in the Emergency Room, or for trauma cases

Level 4: High risk, to be used, for example, during long, fluid intense procedures, surgery, when pathogen resistance is needed or infectious diseases are suspected (non-airborne)

Product Images



FRONT



BACK

Establishment Registration & Device Listing

New Search	Back To Search Results
<p>Establishment: ZHEJIANG BAIYI INTELLIGENT GARMENT CO.,LTD No.1 Fengshan 1 Road Industrial Park ,HangBu Town,KeCheng QuZhou , CN 32400 Registration Number: 3016741301 FEI Number*: 3016741301 Status: Active Date Of Registration Status: 2021</p> <p>Owner/Operator: Zhejiang Baiyi Intelligent Garment Co.,Ltd No.1 Fengshan 1 Road Industrial Park ,Ha QuZhou, CN 32400 Owner/Operator Number: 10063210</p> <p>Official Correspondent: Merlie Porter Three Green Apples, LLC 1002 Strayer Dr. San Jose, CA 95129 Phone: 1-408-7477473</p> <p>US Agent: Merlie Porter Three Green Apples, LLC 1002 Strayer Dr. San Jose , CA US 95129 Phone: 408 7477473 Ext Email: Vicporter@lcloud.Com</p>	
* Firm Establishment Identifier (FEI) should be used for identification of entities within the imports message set	

New Search	Back To Search Results
Proprietary Name:	Isolation Clothing; Single-use Protective Medical Clothing
Classification Name:	NON-SURGICAL ISOLATION GOWN
Product Code:	OEA
Device Class:	1
Regulation Number:	878.4040
Medical Specialty:	General & Plastic Surgery
Registered Establishment Name:	ZHEJIANG BAIYI INTELLIGENT GARMENT CO.,LTD
Registered Establishment Number:	3016741301
Owner/Operator:	Zhejiang Baiyi Intelligent Garment Co.,Ltd
Owner/Operator Number:	10063210
Establishment Operations:	Foreign Exporter; Manufacturer

Product Testing and Certificate



TEST REPORT

(Electronic version)



No: 200137062

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: XEMH-6651-34

扫码下载报告



ISSUE DATE: 2020-06-03

APPLICANT: HUAHAO NONWOVENS CO., LTD
ADDRESS: GAOKE ROAD, LONGGANG CITY, WENZHOU, CHINA

INFORMATION CONFIRMED BY APPLICANT:

SMS nonwoven fabric

Quantity: 1 piece

Type: SMS

DATE RECEIVED/DATE TEST STARTED: 2020-05-25

CONCLUSION:

BURSTING STRENGTH (DRY STATE)	M
BURSTING STRENGTH (WET STATE)	M
BREAKING STRENGTH (DRY STATE)	M
BREAKING STRENGTH (WET STATE)	M
STATIC HYDROSTATIC RESISTANCE	M

NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"---" -NO COMMENT

REMARK:

ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).

COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.

THE EXPERIMENT WAS CARRIED OUT AT No.1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P.R.CHINA.

APPROVED BY:

ZiShan Guo SENIOR ENGINEER

郭子山



总部: 广州市番禺区珠江路1号
花都实验室: 广州市花都区狮岭镇旗岭河滨西路1号

电话: 020-61994598/61994599
电话: 020-37721161

Product Testing and Certificate



CHINA
INSTITUTION
OF
METROLOGY
TESTING
CNAS L10114

TEST REPORT

(Electronic version)

No: 200137062



总部：广州市番禺区珠江路1号
花都实验室：广州市花都区狮岭镇旗岭河滨西路1号

电话：020-61994598/61994599
电话：020-37721151

Product Testing and Certificate



TEST REPORT

(Electronic version)

No:200137062

BURSTING STRENGTH (DRY STATE) (kPa)

(EN ISO 13938-1:1999, DRY STATE, TEST AREA:10cm²)

1# 93.4
2# 123
3# 126
4# 115
5# 98.4

REQUIREMENT

≥40

(SURGICAL GOWN: STANDARD PERFORMANCE
CRITICAL PRODUCT AREA)
(EN 13795-1:2019)

BURSTING STRENGTH (WET STATE) (kPa)

(EN ISO 13938-1:1999, WET STATE, TEST AREA:10cm²)

1# 109
2# 76.6
3# 73.0
4# 84.3
5# 89.2

REQUIREMENT

≥40

(SURGICAL GOWN: STANDARD PERFORMANCE
CRITICAL PRODUCT AREA)
(EN 13795-1:2019)

BREAKING STRENGTH (DRY STATE) (N)

(EN 29073-3:1992, DRY STATE, THE DISTANCE BETWEEN THE CLAMPS:200mm, RATE:100mm/min)
(EN 29073-3:1992 EQV ISO 9073-3:1989)

LENGTH
1# 91.6
2# 90.1
3# 95.0
4# 94.5
5# 76.1
WIDTH
1# 52.9
2# 53.8
3# 52.2
4# 54.9
5# 60.1

REQUIREMENT

≥20

(SURGICAL GOWN: STANDARD PERFORMANCE
CRITICAL PRODUCT AREA)
(EN 13795-1:2019)

BREAKING STRENGTH (WET STATE) (N)

(EN 29073-3:1992, WET STATE, THE DISTANCE BETWEEN THE CLAMPS:200mm, RATE:100mm/min)
(EN 29073-3:1992 EQV ISO 9073-3:1989)

LENGTH
1# 94.6
2# 96.1
3# 98.9
4# 99.8
5# 93.8
WIDTH
1# 54.0
2# 46.5
3# 52.2
4# 49.4
5# 56.7

REQUIREMENT

≥20

(SURGICAL GOWN: STANDARD PERFORMANCE
CRITICAL PRODUCT AREA)
(EN 13795-1:2019)



Page 3 of 4

Product Testing and Certificate



TEST REPORT


(Electronic version)

No:200137062

STATIC HYDROSTATIC RESISTANCE (cmH₂ O)

(EN ISO 811:2018, STEADILY INCREASING WATER PRESSURE:10cmH₂ O/min, TEMPERATURE OF THE WATER:20.0℃, FACE DOWN AND FACE SIDE TESTED)

	REQUIREMENT
1# 39.0	≥20
2# 70.0	(SURGICAL GOWN: STANDARD PERFORMANCE
3# 75.5	CRITICAL PRODUCT AREA)
4# 37.0	(EN 13795-1:2019)
5# 70.5	



—End of Report—

PAGE 4 OF 4



For inquiries or orders contact us here

Sales@360healthsvc.com | 855.360.5460

www.360healthsvc.com

